

Curlin IV Pump

Wash Hands:

With soap and water for at least 15 seconds (sing Happy Birthday) and dry with paper towels.

Installing Batteries:

1. Locate battery door on back of pump.
2. Insert coin into door notch, pushing down on release pin; slide door open in downward motion.
3. Install two “C” batteries into pump and close door by sliding door back into place.

Loading IV tubing:

1. Fully open the door located on the top of the pump.
2. Remove IV tubing from package and locate the **Blue** Tubing-Guide and **Yellow** Flow stop.
3. Insert the **Blue** Tubing-Guide into the receptacle on the right side of the pump (look for the blue arrow).
4. Twist tab off **Yellow** Flow-Stop.
5. Insert **Yellow** Flow-Stop into the receptacle on the left side of the pump (look for the yellow arrow).
6. Close pump door and latch it securely.
7. Remove protective cap from spiked end of tubing.
8. Push spiked end of tubing into IV bag using a twisting/pushing motion.

Yellow Flow-Stop

Blue Tubing-Guide

Curlin IV Pump

Priming IV Tubing/Starting Infusion:

1. Turn pump on (locate **ON/OFF** button on the top left of keypad);
Pump will sound 3 beeps and go through a series of self tests on the display screen.

When pump has completed its self tests the pump then will display the following:

ENTER TO SELECT
PROGRAM
BIOMED SETUP

2. Press the yellow **YES** button located on pump keypad to select “PROGRAM”.

The pump will then display the following:

ENTER TO SELECT
REPEAT RX
NEW PROGRAM

3. Press the yellow **YES** button located on the pump keypad to select “REPEAT RX”.

The pump will then run through its pre-programmed settings and will then display:

RUN TO START
NO TO REVIEW
OPTIONS OR PRIME

4. Press and release the **PRIME/BOLUS** button located at the lower right corner of pump keypad.
5. Press and hold the **PRIME/BOLUS** button until tubing is completely filled with fluid.
6. When priming is complete press **YES** button to exit PRIME screen.
7. Scrub clave with alcohol prep pad (approx. 10-15 seconds)
8. Flush IV catheter using **Saline** (see CNY Infusion Flushing Protocol)
9. Connect IV tubing to clave and then Push the **RUN/PAUSE** button to start infusion.